



### CERTIFICATE OF ANALYSIS

Bioprofile PCO2 Electrode  
Lot Number 16173010

Part Number 23094  
Exp Date 2017-08

|          | Spec      | Pass/Fail |
|----------|-----------|-----------|
| Slope    | 8.7-11.0  | Pass      |
| Std A mV | 0 +/- 150 | Pass      |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

7/12/2016

Date