



CERTIFICATE OF ANALYSIS

Nova PCO2/TCO2 Conditioning Solution

Lot Number 16118038

Ref Number 06857

Exp Date 2017-07

pH	Spec 0.70-1.30	Pass/Fail Pass
----	-------------------	-------------------

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in dark ink, appearing to read "Joyce Amara", written in a cursive style.

Approval

QC Supervisor

Title

5/10/2016

Date