



CERTIFICATE OF ANALYSIS

STP pHox/Basic Cal Cartridge
Lot Number 16018035

Ref Number 23935
Exp Date 2017-07

Cal A L/N	16013014		Cal B L/N	16013019		Cal C L/N	16013016	
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
HCO3 mmol/L	20.1-20.9	Pass	HCO3 mmol/L	40.2-41.8	Pass	pH @ 37C	7.342-7.348	Pass
Na mmol/L	139.3-141.3	Pass						
Cal D L/N	16013017		Reference L/N	16013018				
	Spec	Pass/Fail		Spec	Pass/Fail			
pH @ 37C	6.792-6.798	Pass	KCl	1.8-2.2	Pass			
Na mmol/L	71.5-72.5	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

1/29/2016

Date