



CERTIFICATE OF ANALYSIS

Bioprofile 100/400/Basic/100 Plus Glutamine Sensor
Lot Number 15247036

Part Number 24447
Exp Date 9/30/2017

	Spec*	Pass/Fail
Slope	10.0-60.0	Pass
Biopro Ctl L4	0.85-1.35	Pass
Biopro Ctl L5	4.56-6.06	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Annora", written in a cursive style.

Approval

QC Supervisor

Title

9/21/2015

Date

CONFIDENTIAL