



CERTIFICATE OF ANALYSIS

Bioprofile 100/300/400/Basic/100 Plus Lactate Sensor

Part Number 24449

Lot Number 15212013

Exp Date 7/31/2017

	Spec*	Pass/Fail
Slope	10.0-60.0	Pass
Biopro Ctl L 1	0.89-1.39	Pass
Biopro Ctl L 2	2.55-3.45	Pass
Biopro Ctl L 3	4.25-5.75	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

8/20/2015

Date

CONFIDENTIAL