



### CERTIFICATE OF ANALYSIS

Bioprofile 100/300/400/100 Plus K Sensor

Part Number 24452

Lot Number 16123026

Exp Date 8/31/2017

|          | Spec      | Pass/Fail |
|----------|-----------|-----------|
| Slope    | 10.0-11.2 | Pass      |
| Std A mV | 0 ± 50    | Pass      |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Annala".

Approval

QC Supervisor

Title

5/25/2016

Date