



**CERTIFICATE OF ANALYSIS**

Bioprofile 100/300/400/100 Plus K Sensor  
Lot Number 16168018

Part Number 24452  
Exp Date 9/30/2017

	Spec	Pass/Fail
Slope	10.0-11.2	Pass
Std A mV	0 ± 50	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Approval

QC Supervisor

Title

7/19/2016

Date