



### CERTIFICATE OF ANALYSIS

Bioprofile 100/300/400/100 Plus K Sensor

Part Number 24452

Lot Number 16207012

Exp Date 10/31/2017

	Spec	Pass/Fail
Slope	10.0-11.2	Pass
Std A mV	0 ± 50	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Annala".

Approval

QC Supervisor

Title

8/16/2016

Date