



CERTIFICATE OF ANALYSIS

Bioprofile 100/300 pH Sensor
Lot Number 16055040

Part Number 24453
Exp Date 8/31/2017

	Spec	Pass/Fail
Slope	9.2-11.2	Pass
Std A mV	0±50	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

3/9/2016

Date