



### CERTIFICATE OF ANALYSIS

Bioprofile 100/300 pH Sensor  
Lot Number 16078028

Part Number 24453  
Exp Date 9/30/2017

|          |          |           |
|----------|----------|-----------|
|          | Spec     | Pass/Fail |
| Slope    | 9.2-11.2 | Pass      |
| Std A mV | 0±50     | Pass      |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

3/30/2016

Date