



CERTIFICATE OF ANALYSIS

Bioprofile 100/400/100 Plus Reference Electrode

Part Number 24454

Lot Number 16130042

Exp Date 11/30/2017

	Spec*	Pass/Fail
Na Slope	≥ 9.2	Pass
Std C mV	-80 to +80	Pass
BioPro Ctl L2	123.2-137.2	Pass
K Slope	≥ 10.5	Pass
Std A mV	-50 to +50	Pass
BioPro Ctl L2	3.65-4.35	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with EN ISO 9001:2012, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

5/20/2016

Date