



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100+ Glutamate Membrane

Part Number 24455

Lot Number 16237026

Exp Date 2017-08

	Spec*	Mean Result
Slope	10.0-50.0	19.4
Biopro Ctl L4	0.29-0.79	0.55
Biopro Ctl L5	1.78-2.38	2.05

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

9/13/2016

Date