



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Lactate Membrane
Lot Number 16238012

Part Number 24457
Exp Date 2017-08

	Spec*	Mean Result
Slope	10.0 - 60.0	25.7
Biopro Ctl L1	0.89-1.39	1.07
Biopro Ctl L2	2.55-3.45	3.05
Biopro Ctl L3	4.25-5.75	4.79

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Annala", written in a cursive style.

Approval

QC Supervisor
Title

9/13/2016
Date