



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Glucose Membrane
Lot Number 16179026

Part Number 24458
Exp Date 2017-09

	Spec*	Mean Result
Slope	12.0-60.0	42.5
Biopro Ctl L1	0.75-1.25	1.03
Biopro Ctl L3	12.75-17.25	15.10

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Approval

QC Supervisor

Title

7/14/2016

Date