



CERTIFICATE OF ANALYSIS

STP pHox/pHOx Basic/BiopHOx Air Detector
Lot Number 20203029

Ref Number 21513
Date of Manufacture 2020-07-21
UDI# (01)00385480215138(11)200721(10)20203029

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Leo Bou".

Approval

QC Inspector
Title

8/4/2020
Date