



**CERTIFICATE OF ANALYSIS**

STP pHox/pHOx Basic/BiopHOx Air Detector  
Lot Number 20280013

Ref Number 21513  
Date of Manufacture 2020-10-06  
UDI# (01)00385480215138(11)201006(10)20280013

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in purple ink that reads "Keo Bou".

Approval

QC Inspector  
Title

10/20/2020  
Date