



CERTIFICATE OF ANALYSIS

STP pHOx/pHOx Basic/BiopHOx Air Detector
Lot Number 21362022

Ref Number 21513
Date of Manufacture 2021-12-28
UDI# (01)00385480215138(11)211228(10)21362022

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

1/10/2022
Date