

## **CERTIFICATE OF ANALYSIS**

STP pHOx/pHOx Basic/BiopHOx Air Detector Lot Number 22110001 Ref Number 21513
Date of Manufacture 2022-04-20

<u>Spec</u> <u>Pass/Fail</u> AD1 mV 1400 - 2600 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

5/2/2022

Date