



CERTIFICATE OF ANALYSIS

Bioprofile 100/Basic ADT
Lot Number 15222035

Part Number 25421
Exp Date 8/31/2017

| | | |
|-----------|---------|-----------|
| | Spec | Pass/Fail |
| ADT mV | > 100 | Pass |
| Flow Time | 1.2-2.8 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

8/19/2015

Date

CONFIDENTIAL