



CERTIFICATE OF ANALYSIS

Bioprofile 100/Basic ADT

Part Number 25421

Lot Number 15324005

Exp Date 11/30/2017

	Spec	Pass/Fail
ADT mV	> 100	Pass
Flow Time	1.2-2.8	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in dark ink, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

12/10/2015

Date