



CERTIFICATE OF ANALYSIS

STP pHox Plus Cal Cartridge
Lot Number 16050020

Ref Number 27612
Exp Date 2017-08-12
Date of Manufacture 2016-02-12

Cal A	L/N	16047015		Cal B	L/N	16047016		Cal C	L/N	16047017	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
Na mmol/L		142-144	Pass	HCO ₃ mmol/L		39.2-40.8	Pass	pH @ 37C		7.326-7.332	Pass
HCO ₃ mmol/L		19.6-20.4	Pass					Na mmol/L		72.5-73.5	Pass
K mmol/L		4.03-4.09	Pass					K mmol/L		10.04-10.20	Pass
Cl mmol/L		97.2-98.8	Pass					Cl mmol/L		52.6-53.4	Pass
Ca mmol/L		1.01-1.03	Pass					Glu mg/dL		198-202	Pass
Glu mg/dL		39.6-40.4	Pass								
Cal D	L/N	16043052		Reference	L/N	16043053					
		Spec	Pass/Fail			Spec	Pass/Fail				
pH @ 37C		6.782-6.788	Pass	KCl		1.8-2.2	Pass				
Ca mmol/L		2.01-2.07	Pass								
Glu mg/dL		990-1010	Pass								

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

3/8/2016

Date