



### CERTIFICATE OF ANALYSIS

Nova Cleaning Solution  
Lot Number 16265034

Ref Number 11272  
Exp Date 2017-09-21  
Date of Manufacture 2016-09-21

|       |            |           |
|-------|------------|-----------|
|       | Spec       | Pass/Fail |
| NaOCl | 1.21-1.51% | Pass      |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Annala".

Approval

QC Supervisor

Title

10/5/2016

Date