



CERTIFICATE OF ANALYSIS

Nova Cleaning Solution
Lot Number 16334040

Ref Number 11272
Exp Date 2017-11-29
Date of Manufacture 2016-11-29

	Spec	Pass/Fail
NaOCl	1.21-1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Amara".

Approval

QC Supervisor
Title

1/13/2017
Date