



CERTIFICATE OF ANALYSIS

Bioprofile 300 B Reagent Pack
Lot Number 631259

Part Number 33133
Exp Date 8/31/2017

Std A	L/N	16312059	Std B	L/N	16312059
	Spec	Result		Spec	Result
Na mmol/L	138.6-141.4	Pass	K mmol/L	9.9-10.1	Pass
K mmol/L	3.96-4.04	Pass	NH4 mmol/L	9.9-10.1	Pass
NH4 mmol/L	1.98-2.02	Pass	pH @ 37.0	6.840-6.860	Pass
pH @ 37.0	7.370-7.390	Pass	Gluc g/L	9.5-10.5	Pass
Gluc g/L	1.8-2.2	Pass	Lac g/L	2.75-3.25	Pass
Lac g/L	0.8-1.2	Pass	Ace mmol/L	85-115	Pass
Ace mmol/L	18-22	Pass	Gly g/L	4.5-5.5	Pass
Gly g/L	2.25-2.90	Pass			

Flush	L/N	16312059
	Spec	Result
Aqueous Salt Solution	N/A	Present

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, K, NH4, pH, Gluc, Lac.

Approval

QC Supervisor
Title

12/8/2016
Date