



CERTIFICATE OF ANALYSIS

Bioprofile 300 Reference Electrode
Lot Number 16032015

Part Number 33253

Exp Date

	Spec*	Pass/Fail
Na Slope	≥ 9.8	Pass
Std A mV	-10 to +70	Pass
Set A L3	137.2-143.2	Pass
Cl Slope	≥ 10.5	Pass
Std A mV	+45 to +85	Pass
Set A L3	103.5-111.5	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Approval

QC Supervisor

Title

2/26/2016

Date