



CERTIFICATE OF ANALYSIS

Bioprofile 300/400/100+ Na Sensor
Lot Number 16145031

Part Number 33256
Exp Date 8/31/2017

| | Spec | Pass/Fail |
|----------|------------|-----------|
| Slope | 9.5-11.3 | Pass |
| Std A mV | -10 - +120 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Approval

QC Supervisor

Title

6/7/2016

Date