



CERTIFICATE OF ANALYSIS

Bioprofile 300/400/100+ Na Sensor

Part Number 33256

Lot Number 16187003

Exp Date 10/31/2017

	Spec	Pass/Fail
Slope	9.5-11.3	Pass
Std A mV	-10 - +120	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

8/9/2016

Date