



### CERTIFICATE OF ANALYSIS

Bioprofile 300 Acetate Electrode  
Lot Number 15253009

Part Number 33257  
Exp Date 9/30/2017

	Spec	Pass/Fail
Slope	9.4-11.0	Pass
Std A mV	0 +/- 150	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Annala".

Approval

QC Supervisor  
Title

9/24/2015  
Date

CONFIDENTIAL