



CERTIFICATE OF ANALYSIS

Bioprofile 300/400/100+ Ammonium Sensor
Lot Number 16207015

Part Number 33258
Exp Date 10/31/2017

	Spec	Pass/Fail
Slope	10.5-13.0	Pass
Std J mV	≥ -100	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

A handwritten signature in cursive script, appearing to read "Joyce Annora".

Approval

QC Supervisor
Title

9/19/2016
Date