



CERTIFICATE OF ANALYSIS

Acetate Membrane Kit
Lot Number 15195041

Part Number 33274
Exp Date 7/31/2017

	Spec*	Mean Result
Slope	7.5-16.0	10.2
Biopro Ctl L6	16.5-22.5	18.14
Biopro Ctl L7	80.0-120.0	108.43

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Annala", written in a cursive style.

Approval

QC Supervisor

Title

7/24/2015

Date

CONFIDENTIAL