



CERTIFICATE OF ANALYSIS

Bioprofile 300/400/100+ ADT
Lot Number 15265041

Part Number 33388
Exp Date 9/30/2017

	Spec	Pass/Fail
ADT mV	> 100	Pass
Flow Time	1.5-5.5	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

Approval

QC Supervisor

Title

10/1/2015

Date

CONFIDENTIAL