



# **CERTIFICATE OF ANALYSIS**

STP Preheater Cleaning Agent  
Lot Number 16288032

Ref Number 11802  
Exp Date 2017-10-14  
Date of Manufacture 2016-10-14

	Spec	Pass/Fail
NaOCl	0.43-0.73%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Amara".

Approval

QC Supervisor

Title

10/21/2016

Date