



CERTIFICATE OF ANALYSIS

STP CCX Cal Cartridge 4 w/Creat
Lot Number 109701

Ref Number 37168
Exp Date 2012-09

Cal A L/N	109695		Cal B L/N	109695		Cal C L/N	109695	
	Spec	Pass/Fail		Spec	Pass/Fail		Spec	Pass/Fail
pH @ 37C	7.351-7.357	Pass	pH @ 37C	6.822-6.828	Pass	Na mmol/L	139-141	Pass
Lac mmol/L	13.5-16.5	Pass	Na mmol/L	69.5-70.5	Pass	K mmol/L	3.97-4.03	Pass
			K mmol/L	9.9-10.1	Pass	Cl mmol/L	118.0-120.0	Pass
			Cl mmol/L	39.7-40.3	Pass	BUN mg/dl	9.9-10.1	Pass
			Ca mmol/L	1.97-2.03	Pass	Glu mg/dl	78.8-81.2	Pass
			Mg mmol/L	1.48-1.52	Pass	Lac mmol/L	1.8-2.2	Pass
						Ca mmol/L	0.98-1.02	Pass
						Mg mmol/L	0.49-0.51	Pass
						Creat mg/dl	0.98-1.02	Pass
						HCO3 mmol/L	18.34-19.08	Pass
Cal D L/N	109695		Flush L/N	109695		Reference L/N	109695	
	Spec	Pass/Fail		Spec	Result		Spec	Pass/Fail
BUN mg/dl	49.4-50.6	Pass	Aqueous Salt	N/A	Present	KCl mol/L	1.8-2.2	Pass
Glu mg/dl	197-203	Pass	Solution					
Lac mmol/L	5.5-6.5	Pass						
Creat mg/dl	9.8-10.2	Pass						
HCO3 mmol/L	36.68-38.18	Pass						

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO 13485:2003, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.



Approval

QC Supervisor
Title

10/18/2011
Date

CONFIDENTIAL