



### CERTIFICATE OF ANALYSIS

STP CCX CO-Oximeter Reagent Cartridge  
w/Bilirubin and Depro Solution  
Lot Number 108207

Ref Number 41549  
Exp Date 2012-08

|               |            |         |                |        |         |
|---------------|------------|---------|----------------|--------|---------|
| Dil L/N       | 108207     |         | Flush L/N      | 108207 |         |
|               | Spec       | Result  |                | Spec   | Result  |
| Surfactant    | N/A        | Present | Surfactant     | N/A    | Present |
| Buffers       | N/A        | Present | Preservatives  | N/A    | Present |
| Preservatives | N/A        | Present |                |        |         |
| Depro L/N     | 108206     |         | THb Calibrator | 108206 |         |
|               | Spec       | Result  |                |        |         |
| NaOCl %       | 2.90-3.20% | Pass    | g/dL           | 16.1   |         |
|               |            |         | g/L            | 161    |         |
|               |            |         | mmol/L         | 9.98   |         |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO 13485:2003, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. Total Hemoglobin (THb) is traced by using Cyanmethemoglobin method.

Approval

QC Supervisor

Title

9/22/2011

Date

CONFIDENTIAL