



CERTIFICATE OF ANALYSIS

Bioprofile Flex pH Sensor
Lot Number 21295039

Part Number 42452
Exp Date 2023-10

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	10.0 - 10.8	Pass
Std A mV	0 +/- 110.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Lumanu Ahtu", written in a cursive style.

Approval

QC Inspector

Title

11/2/2021

Date