



CERTIFICATE OF ANALYSIS

Bioprofile Flex Ammonium Sensor
Lot Number 109188

Part Number 42464
Exp Date 2012-09

	Spec*	Pass/Fail
Slope	10.3-12.3	Pass
Std C mV	>/= -100	Pass
BioPro Ctl L1	0.61-1.11	Pass
BioPro Ctl L3	8.24-11.24	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications.

*Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

A handwritten signature in dark ink, appearing to read "Joyce Amara", written in a cursive style.

Approval

QC Supervisor

Title

10/4/2011

Date

CONFIDENTIAL