



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamine Membrane

Lot Number 21285010

Part Number 42486

Exp Date 2022-04

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	20.99
Biopro Ctl L4	0.69 - 1.19	1.02
Biopro Ctl L5	4.36 - 5.86	4.86

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector

Title

10/28/2021

Date