



CERTIFICATE OF ANALYSIS

Nova 1/5/10/11/13 Microsample Std
Lot Number 108198

Ref Number 13984
Exp Date 2012-08

	Spec	Pass/Fail
Na mmol/L	27.6-28.4	Pass
K mmol/L	0.79-0.81	Pass
Cl mmol/L	22.6-23.3	Pass
Li mmol/L	0.098-0.102	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO 13485:2003, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in dark ink, appearing to read "Joyce Anzara".

Approval

QC Supervisor
Title

9/1/2011
Date

CONFIDENTIAL