



CERTIFICATE OF ANALYSIS

PCO2 Internal Filling Solution
Lot Number 108721

Ref Number 06553
Exp Date 2012-08

	Spec	Pass/Fail
NaCl mmol/L	145.5-154.5	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and ISO 13485:2003, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Annora".

Approval

QC Supervisor
Title

10/5/2011
Date

CONFIDENTIAL