



**CERTIFICATE OF ANALYSIS**

Nova Control Levels 1 and 2

Ref Number 18655

Level 1 L/N	15229034		Level 2 L/N	15229035	
Level 1 Exp Date	2017-08		Level 2 Exp Date	2017-08	
	Spec	Pass/Fail		Spec	Pass/Fail
Na+ mmol/L	132-156	Pass	Na+ mmol/L	110-129	Pass
K+ mmol/L	3.2-4.0	Pass	K+ mmol/L	5.6-6.8	Pass
Cl- mmol/L	95-114	Pass	Cl- mmol/L	76-92	Pass
TCO2 mmol/L	19-30	Pass	TCO2 mmol/L	10-18	Pass
Glu mg/dL	69-91	Pass	Glu mg/dL	233-309	Pass
BUN mg/dL	15-25	Pass	BUN mg/dL	48-68	Pass
TCa mg/dL	5.04-7.24	Pass	TCa mg/dL	7.74-10.91	Pass
Ca++ mg/dL	3.96-5.72	Pass	Ca++ mg/dL	6.49-8.85	Pass
Mg++ mg/dL	0.98-1.74	Pass	Mg++ mg/dL	3.38-4.65	Pass
Li+ mmol/L	0.56-0.87	Pass	Li+ mmol/L	1.84-2.56	Pass
pH	7.30-7.40	Pass	pH	7.55-7.60	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

9/24/2015

Date

**CONFIDENTIAL**