



CERTIFICATE OF ANALYSIS

Performance Check Solution
Lot Number 22213035

Ref Number 17457
Exp Date 2024-02-01
Date of Manufacture 2022-08-01
UDI# (01)00385480174572(11)220801(17)240201(10)22213035

	<u>Spec</u>	<u>Pass/Fail</u>
Glu mg/dL	689.5 - 710.5	Pass
BUN mg/dL	97.0 - 103.0	Pass
Creatinine mg/dL	1.54 - 1.66	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

8/18/2022

Date