



CERTIFICATE OF ANALYSIS

Linearity Std Set A Level 6
Lot Number 16014021

Ref Number 14071
Exp Date 2017-07

	Spec	Pass/Fail
Na mmol/L	98.0-102.0	Pass
K+ mmol/L	98.0-102.0	Pass
Cl- mmol/L	196.0-204.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

A handwritten signature in cursive script, appearing to read "Joyce Ansara".

Approval

QC Supervisor

Title

1/26/2016

Date