



CERTIFICATE OF ANALYSIS

Bioprofile Flex Linearity Solution HIGH
Lot Number 16132051

Catalog Number 45279
Exp Date 2017-11

	Spec	Pass/Fail
Gln mmol/L	4.60-5.40	Pass
Glu mmol/L	4.60-5.40	Pass
Gluc g/L	13.50-16.50	Pass
Lac g/L	4.60-5.40	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM materials: Gluc, Lac

A handwritten signature in cursive script, appearing to read "Joyce Anzara".

Approval

QC Supervisor

Title

6/15/2016

Date