



## CERTIFICATE OF ANALYSIS

BioPro Flex Deproteinizing Solution  
Lot Number 108021

Ref Number 46850  
Exp Date 2012-08

|         | Spec     | Pass/Fail |
|---------|----------|-----------|
| NaOCl % | 0.7-0.9  | Pass      |
| pH      | 9.0-11.0 | Pass      |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP) , and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

A handwritten signature in dark ink, appearing to read "Joyce Ansara", written in a cursive style.

Approval

QC Supervisor

Title

9/1/2011

Date

CONFIDENTIAL