



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Chemistry Controls

Ref Number 49184

Lot Number 16307103

Exp Date 2017-10-26
Date of Manufacture 2016-10-26

Level 4 L/N	16300032		Level 5 L/N	16300033	
Level 4 Exp Date	2017-10-26		Level 5 Exp Date	2017-10-26	
	Spec	Pass/Fail		Spec	Pass/Fail
Na+ mmol/L	134.8-137.6	Pass	Na+ mmol/L	110.6-112.8	Pass
K+ mmol/L	3.87-3.95	Pass	K+ mmol/L	6.19-6.31	Pass
Cl- mmol/L	122.4-124.8	Pass	Cl- mmol/L	96.5-98.5	Pass
Ca++ mmol/L	0.97-0.99	Pass	Ca++ mmol/L	1.48-1.50	Pass
Mg++ mmol/L	0.51-0.63	Pass	Mg++ mmol/L	0.95-1.07	Pass
Glu mg/dL	76-80	Pass	Glu mg/dL	270-282	Pass
Lac mmol/L	1.5-2.1	Pass	Lac mmol/L	6.5-7.1	Pass
BUN mg/dL	18.5-19.5	Pass	BUN mg/dL	56.6-59.5	Pass
Creatinine mg/dL	1.08-1.12	Pass	Creatinine mg/dL	6.27-6.53	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

2/8/2017
Date