



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1
Lot Number 0421043301

Ref Number 41741
Exp Date 2023-08-12
Date of Manufacture 2021-02-12
UDI (01)00385480417419(11)210212(17)230812(10)0421043301

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

3/1/2021
Date