



CERTIFICATE OF ANALYSIS

Bioprofile pH Linearity Solution HIGH
Lot Number 108750

Catalog Number 49938
Exp Date 2012-08

Spec	Pass/Fail
pH 7.725-7.925	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM materials: pH

A handwritten signature in cursive script, appearing to read "Joyce Annala".

Approval

QC Supervisor

Title

10/17/2011

Date

CONFIDENTIAL