



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16254001

Ref Number 52427
Exp Date 2017-08-17
Date of Manufacture 2016-08-17

Cal A L/N 16230048

Cal B L/N 16231091

Flush L/N 16231092

	Spec	Pass/Fail
pH	7.340	N/A
pCO2 mmHg	25.1	N/A
Na mmol/L	132.1-133.9	Pass
K mmol/L	3.97-4.03	Pass
Cl mmol/L	105.2-106.8	Pass
iCa mmol/L	1.08-1.12	Pass
Glu mg/dL	79-81	Pass
Lac mmol/L	1.8-2.2	Pass

	Spec	Pass/Fail
pH	6.837-6.843	Pass
Na mmol/L	68.0-69.0	Pass
K mmol/L	9.92-10.08	Pass
Cl mmol/L	45.6-46.4	Pass
iCa mmol/L	2.17-2.23	Pass
Glu mg/dL	197-203	Pass
Lac mmol/L	9.5-10.5	Pass

	Spec	Pass/Fail
pCO2 mmHg	49.8	N/A
pO2 mmHg	104.2	N/A

Reference L/N 16231007

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

10/5/2016

Date