



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16354046

Ref Number 52427
Exp Date 2017-10-26
Date of Manufacture 2016-10-26

Cal A L/N 16300042

Cal B L/N 16300034

Flush L/N 16316042

	Spec	Pass/Fail
pH	7.353	N/A
pCO2 mmHg	24.6	N/A
Na mmol/L	132.1-133.9	Pass
K mmol/L	3.97-4.03	Pass
Cl mmol/L	105.2-106.8	Pass
iCa mmol/L	1.08-1.12	Pass
Glu mg/dL	79-81	Pass
Lac mmol/L	1.8-2.2	Pass

	Spec	Pass/Fail
pH	6.837-6.843	Pass
Na mmol/L	68.0-69.0	Pass
K mmol/L	9.92-10.08	Pass
Cl mmol/L	45.6-46.4	Pass
iCa mmol/L	2.17-2.23	Pass
Glu mg/dL	197-203	Pass
Lac mmol/L	9.5-10.5	Pass

	Spec	Pass/Fail
pCO2 mmHg	51.6	N/A
pO2 mmHg	101.8	N/A

Reference L/N 16300021

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Approval

QC Supervisor

Title

1/10/2017

Date