



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge
Lot Number 16323025

Ref Number 53105
Exp Date 2017-09-28
Date of manufacture 2016-09-28

Cal A	L/N	16272031		Cal B	L/N	16272033		Flush	L/N	16301037	
		Spec	Pass/Fail			Spec	Pass/Fail			Spec	Pass/Fail
	pH	7.339	N/A		pH	6.837-6.843	Pass		pCO2 mmHg	51.2	N/A
pCO2 mmHg		25.6	N/A	Na mmol/L		68.0-69.0	Pass		pO2 mmHg	102.0	N/A
Na mmol/L	132.1-133.9		Pass	K mmol/L	9.92-10.08		Pass				
K mmol/L	3.97-4.03		Pass	Cl mmol/L	45.6-46.4		Pass				
Cl mmol/L	105.2-106.8		Pass	iCa mmol/L	2.17-2.23		Pass				
iCa mmol/L	1.08-1.12		Pass	Glu mg/dL	197-203		Pass				
Glu mg/dL	79-81		Pass	Lac mmol/L	9.5-10.5		Pass				
Lac mmol/L	1.8-2.2		Pass								

Reference L/N 16272034

	Spec	Pass/Fail
KCl	1.8-2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2012, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.


Approval

QC Supervisor
Title

1/13/2017
Date